

分銅校正の拡張不確かさ (k=2)

恒久的施設で行う校正

校正範囲	F1準拠	F2準拠	M1/M2準拠
1000 kg			25 g
500 kg			13 g
20 kg	22 mg	44 mg	0.12 g
10 kg	13 mg	23 mg	64 mg
5 kg	7.3 mg	13 mg	33 mg
2 kg	2.8 mg	4.9 mg	13 mg
1 kg	1.6 mg	2.6 mg	6.5 mg
500 g	0.71 mg	1.3 mg	3.3 mg
200 g	0.33 mg	0.52 mg	1.3 mg
100 g	0.26 mg	0.35 mg	0.71 mg
50 g	0.064 mg	0.13 mg	0.39 mg
20 g	0.053 mg	0.11 mg	0.32 mg
10 g	0.039 mg	0.083 mg	0.24 mg
5 g	0.032 mg	0.067 mg	0.20 mg
2 g	0.026 mg	0.051 mg	0.16 mg
1 g	0.021 mg	0.042 mg	0.12 mg
500 mg	0.017 mg	0.034 mg	0.098 mg
200 mg	0.015 mg	0.027 mg	0.079 mg
100 mg	0.012 mg	0.022 mg	0.063 mg
50 mg	0.0090 mg	0.018 mg	0.048 mg
20 mg	0.0082 mg	0.015 mg	0.040 mg
10 mg	0.0071 mg	0.012 mg	0.032 mg
5 mg	0.0060 mg	0.0099 mg	0.025 mg
2 mg	0.0058 mg	0.0097 mg	0.025 mg
1 mg	0.0055 mg	0.0094 mg	0.024 mg